SIEMENS

IMMULITE 2000/XPi 3gAllergy Specific IgE Birch Pollen Major Allergen, nBet v 1 (Betula verrucosa, Code A89L2)

www.siemens.com/allergy

Background

IMM Birch (*Betula verrucosa*) pollen is a major cause of allergy IMM in the northern parts of Europe and America that attributed IMM to about 22% of allergic individuals suffering from pollinosis.1 IMM Six birch pollen allergens have already been identified IMM (Bet v 1, Bet v 2, Bet v 3, Bet v 4, Bet v 6, Bet v 7). Bet v 1, IMM a major allergenic protein of birch pollen, is a 17-kD protein IMM consisting of several isoallergens² and recognized by IgE IMM antibodies from almost all birch pollen-allergic patients.³ IMM Oral allergy syndrome (OAS) as a result of primary IMM sensitization to Bet v 1 has been previously reported.



Testing Algorithm⁴⁻⁶



Biochemical Characteristics

Native Pru p 3 (nPru p 3) protein was purified by conventional methods from *Prunus persica*.



Figure 1. Coomassie Blue stained gel.

Clinical Performance

Clinical performance was demonstrated by testing serum samples from clinically diagnosed atopic and apparently healthy individuals against the nBet v 1 specific allergen. The results were obtained using the IMMULITE[®] 2000 3gAllergy[™] Specific IgE assay. Overall agreement, sensitivity, and specificity are presented in the cross table next page.

Answers for life.

Allergen: nBet v 1

IMMULITE 2000				
	Clinical	Normal	Total	
Positive (≥0.10 kU/L)	37	5	42	
Negative	6	95	101	
Total	43	100	143	
Sensitivity (95% Confidence Interval)	Specificity (95% Confidence Interval)		Overall Agreement	
86% (76% to 96%)	95% (91% to 99%)		92%	

Additional clinical performance of the nBet v 1 specific allergen was demonstrated in comparison to the whole extract allergen T3 (Birch). A total of 143 samples were tested with A89 and T3. The results are presented below.

Allergen: nBet v 1

IMMULITE 2000					
	T3 (Ref. M				
A89	42	0	Positive		
(Test Method)	11	90	Negative		
	Positive	Negative			

N=143 Overall percent agreement = 92% Positive percent agreement = 79% Negative percent agreement = 100%

Analytical Performance

Precision: The average within-run and total precision using three samples and three lots of nBet v 1 allergen was 3.58% and 4.13%, respectively.

Linearity: Two samples were diluted in 2-fold serial dilutions. The undiluted (neat) and the diluted samples were assayed in two replicates and the observed value was reported based on the average of the two replicates. Comparisons of the observed to expected values were used to demonstrate linearity at concentrations within the assay limits.

Observed = 0.994 (Expected) + 0.1429

Slope (95% Confidence Interval) = 0.994 (0.972 to 1.016)

Identity Testing

Identity of nBet v 1 allergen was verified through competitive inhibition testing using a single serum sample. A negative sample was used to measure the background response. The percentage inhibitions are represented in the graph below showing correlation to increasing inhibitor concentrations.



References:

- 1. Horak F, Jager S. Die Erreger des Heufiebers. 19 pp. Verlag Urban
- & Schwarzenverg, Munich-Vienna-Baltimore, 1979. 2. Swoboda I, et al. J Biol Chem. 1995;270:2607-13.
- Swoboda I, et al. J Biol Cheffi. 1995;270:2607-15.
 Ipsen H, Lowenstein H. J. Allergy Clin Immunol. 1983;72:150-9.
- 4. Menz G, et al. Clin Exp Allergy. 1996;26(1):50-60.
- 4. Menz G, et al. Chin Exp Anergy. 1996;26(1):50-60.
- 5. Kazemi-Shirazi L, et al. J Allergy Clin Immunol. 2000;105(1 Pt 1):116-25.
- 6. Moverare R, et al. Int Arch Allergy Immunol. 2002;128(4):325-35.

Siemens Healthcare Diagnostics, a global leader in clinical diagnostics, provides healthcare professionals in hospital, reference, and physician office laboratories and point-of-care settings with the vital information required to accurately diagnose, treat, and monitor patients. Our innovative portfolio of performance-driven solutions and personalized customer care combine to streamline workflow, enhance operational efficiency, and support improved patient outcomes.

3gAllergy, IMMULITE, and all associated marks are trademarks of Siemens Healthcare Diagnostics Inc. All other trademarks and brands are the property of their respective owners. Product availability may vary from country to country and is subject to varying regulatory requirements. Please contact your local representative for availability.

Global Siemens Headquarters

Siemens AG Wittelsbacherplatz 2 80333 Muenchen Germany

Global Siemens Healthcare

Headquarters Siemens AG Healthcare Sector Henkestrasse 127 91052 Erlangen, Germany Phone: +49 9131 84 - 0 www.siemens.com/healthcare

Global Division

Siemens Healthcare Diagnostics Inc. 511 Benedict Avenue Tarrytown, NY 10591-5005 USA www.siemens.com/diagnostics

Order No. A91DX-CAI-121339-GC1-4A00 01-2013 | All rights reserved © 2013 Siemens Healthcare Diagnostics Inc.